



ACTUAL SIZE

WARRANTY: While Filtrertek Inc. makes every reasonable effort to ensure the accuracy and efficiency of its design and manufacturing processes, the Customer bears the final responsibility for determining the suitability, safety, effectiveness, marketability and over-all fitness of the product for any intended purpose.

UNITS ARE IN: INCHES File Server No. 0064260.PRT

This print and all information thereon is confidential and remains the property of Filtrertek
Copyright © 1992 Filtrertek

TW Medical 11411 PRICE ROAD, HEBRON ILLINOIS 60034

FRAME SPEC. CYROLITE G-20 HIFLO NATURAL	PART NAME 17MM TP - 100% E.D.
SCREEN SPEC. .45µ GORE/PET L32426	SHT NO 1 of 1 DATE 12/15/92 DRAWN SGM
	XP NO ---- SCALE 4:1 CHECK CB
CUST. PART NO. --	REV. --
CUST. NAME --	DATE ---
TOLERANCE ON DIMENSION UNLESS OTHERWISE SPECIFIED:	
ENGLISH: TWO PLACE DIM ±.01 THREE PLACE DIM: ±.005	
METRIC: ONE PLACE DIM ±0.3 TWO PLACE DIM: ±0.13	
ANGULAR: ±2°	
PART NUMBER 64260	