




ACTUAL SIZE

NOTES:

- 1) FILTER AND HOUSING TO BE LEAK FREE WHEN TESTED PER A-1792-252. (CUSTOMER TEST REF.)
- 2) FLOW RATE TO BE 4 LITERS PER MINUTE MINIMUM WHEN TESTED PER A-1792-252. (CUSTOMER TEST REF.)
- 3) REFERENCE FILTERTEK PART #61310. THIS PART IS IDENTICAL BUT WITH CUSTOM FILTER.

WARRANTY: While Filtertek Inc. makes every reasonable effort to ensure the accuracy and efficiency of its design and manufacturing processes, the Customer bears the final responsibility for determining the suitability, safety, effectiveness, marketability and overall fitness of the product for any intended purpose.					
UNITS ARE IN: INCHES			File Server No. 0085280.PRT		
This print and all information thereon is confidential and remains the property of Filtertek Copyright © 1992 Filtertek					
			21411 PRICE ROAD, HEBRON ILLINOIS 60034		
FRAME SPEC: FINA CY80 8-20 HEILD NATURAL			PART NAME: 17 MM INLINE FILTER		
SCREEN SPEC: PAUL BRPAB T500A30 1ST			SHV NO: 141	DATE: 4-8-93	OSAWN: SM
REV: A			DATE: 3/18/93	SCALE: 8:1	CHECK
CUST. PART NO: B-1792-253			TOLERANCE ON DIMENSION UNLESS OTHERWISE SPECIFIED:		
CUST. NAME: SHERWOOD MEDICAL			ENGLISH: TWO PLACE DIM ±.03 THREE PLACE DIM: ±.010		
			METRIC: ONE PLACE DIM ±0.75 TWO PLACE DIM: ±0.25		
			ANGULAR: 11/2°		
			65280		