



SECTION A-A



DRAWING NO.  
0056080-00

# TW Medical

PART NO.	PORT I.D.	HYDROPHILIC SCREEN
0069400	.090"	0.2u PES
0069410	.090"	1.2u PES

WARRANTY: While Filtertek Inc. makes every reasonable effort to ensure the accuracy and efficiency of its design and manufacturing processes, the Customer bears the final responsibility for determining the suitability, safety, effectiveness, marketability and over-all fitness of the product for any intended purpose.

UNITS ARE IN: INCHES | File Server No. 0056080.PRT

This print and all information thereon is confidential and remains the property of Filtertek  
Copyright © 2003 Filtertek

**TW Medical**

11411 PRICE ROAD, HEBRON ILLINOIS 60034

FRAME SPEC.		ACRYLIC		PART NAME	MICRO INFUSION FILTER ASSEMBLY	
SCREEN SPEC.	SEE CHART	SHT No.	-of-	DATE	DRAWN TO	
		XP No.	----	SCALE	4:1	CHECK
CUST. PART NO.	REV.	TOLERANCE ON DIMENSION UNLESS OTHERWISE SPECIFIED:				DRAWING NO.
--	--	ENGLISH: TWO PLACE DIM ±.01 THREE PLACE DIM: ±.005				0056080-00
CUST. NAME	DATE	METRIC: ONE PLACE DIM ±0.3 TWO PLACE DIM: ±0.13				
--	---	ANGULAR: ±2°				